

## INSTRUCTIONS FOR TYING BALLOONS

- Use surgical silk suture, size 00 or 000.
- Insert distal tip of catheter into balloon.
- Wrap suture around each end of the balloon onto the catheter, ensuring the suture is flat and does not overlap onto itself.
- Tie balloon onto catheter tight enough so that there are no leaks but not too tight that the inside lumens collapse.
- Make a surgical knot.
- Use any medical grade latex glue or cyanoacrylate to seal the knots.

**Caution:** Cidex-OPA (manufactured by Johnson & Johnson) is NOT compatible with these balloons.

**Note:** **Balloon shelf life:** 2 years  
**Storage Requirements:** Store at room temperature (20° to 25°C or 68°-77°F). Do not expose balloons to direct light.

Mui Scientific is not liable for any damage to the catheter, or harm to patients or personnel, caused by improper use of these procedures.



Edit Date: August 25, 2016



**Mui Scientific**

145 Traders Blvd. E., Unit #34, Mississauga, ON, Canada, L4Z 3L3  
Tel: (905) 890-5525 Toll Free: 1-800-303-6611 Fax: (905) 890-3523  
Website: Muiscientific.com Email: mail@muiscientific.com

## INSTRUCTIONS FOR TYING BALLOONS

- Use surgical silk suture, size 00 or 000.
- Insert distal tip of catheter into balloon.
- Wrap suture around each end of the balloon onto the catheter, ensuring the suture is flat and does not overlap onto itself.
- Tie balloon onto catheter tight enough so that there are no leaks but not too tight that the inside lumens collapse.
- Make a surgical knot.
- Use any medical grade latex glue or cyanoacrylate to seal the knots.

**Caution:** Cidex-OPA (manufactured by Johnson & Johnson) is NOT compatible with these balloons.

**Note:** **Balloon shelf life:** 2 years  
**Storage Requirements:** Store at room temperature (20° to 25°C or 68°-77°F). Do not expose balloons to direct light.

Mui Scientific is not liable for any damage to the catheter, or harm to patients or personnel, caused by improper use of these procedures.



Edit Date: August 25, 2016



**Mui Scientific**

145 Traders Blvd. E., Unit #34, Mississauga, ON, Canada, L4Z 3L3  
Tel: (905) 890-5525 Toll Free: 1-800-303-6611 Fax: (905) 890-3523  
Website: Muiscientific.com Email: mail@muiscientific.com