



Mui Scientific

Division of  
H&A Mui  
Enterprises Inc.

145 Traders Blvd. E., Unit #34, Mississauga, ON, Canada, L4Z 3L3  
Tel: (905) 890-5525 Toll Free: 1-800-303-6611 Fax: (905) 890-3523  
Website: [www.muiscientific.com](http://www.muiscientific.com) Email: [mail@muiscientific.com](mailto:mail@muiscientific.com)

**INSTRUCTION FOR USE FOR  
SINGLE-USE MANOMETRIC WATER-PERFUSED AND/OR BAROSTAT CATHETERS**

- This catheter is designed for **Single-Use Only**. Attempts to reprocess, resterilize, and/or re-use may lead to failure of the device and/or cross-contamination.
- Inspect catheter for any evidence of mechanical damage or imperfections. **DO NOT USE IF DAMAGED.**
- This catheter must only be used under medical supervision by personnel who have been adequately trained in safe gastro-intestinal intubation, and in the performance of perfused manometric measurements and/or barostat measurements.
- Post-procedure manometric catheters are contaminated with body fluid. Immediately dispose of device to prevent re-use.
- **MUI SCIENTIFIC IS NOT LIABLE FOR ANY DAMAGE TO THE CATHETER, OR HARM TO THE PATIENTS OR PERSONNEL, CAUSED BY IMPROPER USE.**

CE  
0120

#  
#  
#  
#  
#  
#

Edit date: July 7, 2016



Mui Scientific

Division of  
H&A Mui  
Enterprises Inc.

145 Traders Blvd. E., Unit #34, Mississauga, ON, Canada, L4Z 3L3  
Tel: (905) 890-5525 Toll Free: 1-800-303-6611 Fax: (905) 890-3523  
Website: [www.muiscientific.com](http://www.muiscientific.com) Email: [mail@muiscientific.com](mailto:mail@muiscientific.com)

#  
#

**INSTRUCTION FOR USE FOR  
SINGLE-USE MANOMETRIC WATER-PERFUSED AND/OR BAROSTAT CATHETERS**

- This catheter is designed for **Single-Use Only**. Attempts to reprocess, resterilize, and/or re-use may lead to failure of the device and/or cross-contamination.
- Inspect catheter for any evidence of mechanical damage or imperfections. **DO NOT USE IF DAMAGED.**
- This catheter must only be used under medical supervision by personnel who have been adequately trained in safe gastro-intestinal intubation, and in the performance of perfused manometric measurements and/or barostat measurements.
- Post-procedure manometric catheters are contaminated with body fluid. Immediately dispose of device to prevent re-use.
- **MUI SCIENTIFIC IS NOT LIABLE FOR ANY DAMAGE TO THE CATHETER, OR HARM TO PATIENTS OR PERSONNEL, CAUSED BY IMPROPER USE.**

CE  
0120

Edit date: July 7, 2016